



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certification Accreditation
(Certificate No : L3319-230215)

This is to certify that

Sherex Taiwan Co.
Mechanical Lab

No. 201, Sandong Rd, Chungli Dist, Taoyuan City, Taiwan. (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2017 ; CNS 17025:2018

Accreditation Number : 3319

Originally Accredited : March 30, 2017

Effective Period : March 30, 2023 to March 29, 2026

Accredited Scope : Testing Field, see described in the Appendix



Scan to verify

Ching-Chang Lien

Ching-Chang Lien
President, Taiwan Accreditation Foundation
February 15, 2023

Accreditation Number : 3319

Laboratory Head : KUO, Gerry

▀ 01. 01 Metals and Alloys Products

Metallic Materials

M101 Rockwell Hardness

ASTM E18

(45 to 100) HRBW

(20 to 65) HRC

Approval Signatory: KUO, Gerry

M103 Vickers Hardness

ASTM E384

ASTM E92

<240 HV0.2

(240 to 600) HV0.2

<240 HV0.3

(240 to 600) HV0.3

<240 HV0.5

(240 to 600) HV0.5

<240 HV10

(240 to 600) HV10

Approval Signatory: KUO, Gerry

▀ 01. 03 Metals and Alloys Products

Fasteners

M002 Axial Tensile Strength

ASTM F606/F606M

ISO 898-1

SAE J429

(2000 to 18000) kgf

(19.61 to 176) kN

Approval Signatory: KUO, Gerry

M009 Plating Thickness (X-ray Method)

ASTM B568

Zn/Fe: (4 to 26) μm

ZnNi/Fe: (5 to 17) μm

Approval Signatory: KUO, Gerry

P2, total 4 pages

The Appendix forms an integral part of this Certificate, which shall be invalid when use without the Appendix



M012 Salt Spray Test
ASTM B117

Approval Signatory: KUO, Gerry

M101 Rockwell Hardness
ASTM E18
ASTM F606/F606M
ISO 898-1
ISO 898-2
(45 to 100) HRBW
(20 to 65) HRC

Approval Signatory: KUO, Gerry

M103 Vickers Hardness
ASTM E384
ASTM E92
ASTM F606/F606M
ISO 898-1
ISO 898-2
<240 HV0.2
(240 to 600) HV0.2
<240 HV0.3
(240 to 600) HV0.3
<240 HV0.5
(240 to 600) HV0.5
<240 HV10
(240 to 600) HV10

Approval Signatory: KUO, Gerry

▀ 01. 03 Metals and Alloys Products

Internal Thread Fasteners
M014 Proof Load (Tension Method)
ASTM F606/606M
ISO 898-2
SAE J995
(2000 to 18000) kgf
(19.61 to 176) kN

Approval Signatory: KUO, Gerry



01.03 Metals and Alloys Products
External Thread Fasteners
M014 Proof Load (Length Measurement Method)
ASTM F606/606M
ISO 898-1
SAE J429
(2000 to 18000) kgf
(19.61 to 176) kN

Approval Signatory: KUO, Gerry

(Null below)

